

GE Infrastructure
Sensing

DECLARATION
OF
CONFORMITY

We,

Panometrics Limited
Shannon Industrial Estate
Shannon, County Clare
Ireland

declare under our sole responsibility that the

**Moisture Monitor Series 35IS Portable Analyzer
PM880 Portable Moisture Monitor**


to which this declaration relates, are in conformity with the following standards:

- EN 50014:1997+A1+A2:1999
- EN 50020:1994
- EN50284:1999
- II 1 G EEx ia IIC T4
PM880: BASEEFA02ATEX0191
Baseefa (2001) Ltd/EECS, Buxton SK17 9JN, UK
MMS 35IS: ISSeP01ATEX035 X; ISSeP, B7340 Colfontaine, Belgium
- EN 61326:1998, Class A, Annex C, Continuous Unmonitored Operation
(for EN 61000-4-3, the MMS 35IS meets performance Criteria A and, in a limited number of frequencies, performance Criteria B per EN 61326.)

following the provisions of the 89/336/EEC EMC Directive and the 94/9/EC ATEX Directive.

The units listed above and any sensors and ancillary sample handling systems supplied with them do not bear CE marking for the Pressure Equipment Directive, as they are supplied in accordance with Article 3, Section 3 (sound engineering practices and codes of good workmanship) of the Pressure Equipment Directive 97/23/EC for DN<25.

Shannon - July 1, 2003



Mr. James Gibson
GENERAL MANAGER



CERT-DOC-H3



August 2004)