

RKI Instruments PS 2 Combustible Gas Monitor - Required Information

When placing your order the RKI Instruments PS 2 Combustible Gas Monitor, we require additional information to process this order:

1. What is the target gas or gas you would like us to calibrate to?

Partial List Of Detectable Gases	
Many other gases available. Contact RKI for gases not listed	
Ethyl Alcohol	C_2H_5OH
Hexane	C_6H_{14}
Hydrogen	H_2
Iso-Butane	$i-C_4H_{10}$
Isopropyl Alcohol	$(CH_3)_2CHOH$
Methane	CH_4
Propane	C_3H_8

2. What would you like the Alarm 1 and Alarm 2 set point to be?

NOTE: Alarm 1 and Alarm 2 set point can be the same LEL levels available for many gases. Typically set to 10% LEL for alarm 1, and 25% or 30% LEL for alarm 2 (Highest possible setting is 30% LEL).

PLEASE INCLUDE THIS INFORMATION IN THE COMMENTS SECTION WHEN YOU PLACE YOUR ON-LINE ORDER.