



Bluetooth®/IRDA printer

0554 0620/ 0554 0621

Instruction manual



1 Contents

1	Contents	3
2	Safety and the environment	4
3	Specifications	4
	3.1. Use.....	4
	3.2. Technical data.....	4
4	Product description	10
	4.1. Overview.....	10
	4.2. Operating mode indicator.....	11
	4.3. Control key functions.....	11
5	First steps	12
	5.1. Charging the rechargeable battery.....	12
	5.2. Inserting paper.....	12
6	Using the product	13

2 Safety and the environment

- > Dispose of faulty rechargeable batteries/spent batteries in accordance with the valid legal specifications.
- > At the end of its useful life, send the product to the separate collection for electric and electronic devices (observe local regulations) or return the product to Testo for disposal.

3 Specifications

3.1. Use

The testo Bluetooth®/IRDA printer is used to produce report printouts, in conjunction with compatible devices:
E.g. testo 330 (0632 3306 - 3307 from firmware 2.02 onwards),
testo 320 (from firmware 1.06 onwards), testo 324 (from firmware 1.08 onwards), testo 330i (0554 0621 only)



The use of the wireless module is subject to the regulations and stipulations of the respective country of use, and the module may only be used in countries for which a country certification has been granted. The user and every owner has the obligation to adhere to these regulations and prerequisites for use, and acknowledges that the re-sale, export, import etc. in particular in countries without wireless permits, is his responsibility.

3.2. Technical data

General data

Feature	Values
Model	Thermal printer with Bluetooth® and IrDA interface
Printing speed	30 mm/s
Operating temperature	0 to +50 °C / 32 to 122 °F
Storage temperature	-20 to +50 °C / -4 to 122 °F
Dimensions (LxWxH)	150 mm x 80 mm x 41 mm

Feature	Values
Weight	approx. 400 g (including rechargeable battery pack, paper roll, mains unit)
EU Directive	2014/30/EC
Power supply	Lithium-Ion rechargeable battery pack, 2600 mAh, 3.7 V
Thermal paper roll	56.5 mm x 34 mm x 12 mm
Rechargeable battery storage temperature	±0 to 35 °C / 32 to 95 °F
Rechargeable battery charge time	approx. 5-6 h
Rechargeable battery life	approx. 50 m report printouts
Bluetooth® (optional)	Range < 10 m

Bluetooth® data 0554 0620





Feature	Values
Type designation	Bluegiga WT12
Specification	2.0
Radio class	Class 2
Coverage	< 10 m / < 32.8 ft.
Product note	WT12
Identification	B011198
Company	10274
Certification	EU countries: Belgium (BE), Bulgaria (BG), Denmark (DK), Germany (DE), Estonia (EE), Finland (FI), France (FR), Greece (GR), Ireland (IE), Italy (IT), Latvia (LV), Lithuania (LT), Luxembourg (LU), Malta (MT), Netherlands (NL), Austria (AT), Poland (PL), Portugal (PT), Romania (RO), Sweden (SE), Slovakia (SK), Slovenia (SI), Spain (ES), Czech Republic (CZ), Hungary (HU), United

Feature	Values
	<p data-bbox="460 193 893 220">Kingdom (GB), Republic of Cyprus (CY).</p> <hr/> <div data-bbox="473 244 527 304" style="display: inline-block; border: 1px solid black; padding: 2px; text-align: center; width: 30px; height: 30px; line-height: 30px;">i</div> <p data-bbox="553 237 930 352">You can find the EU declaration of conformity on the Testo website, www.testo.com, under the product-specific downloads.</p> <hr/> <p data-bbox="460 395 647 419">EFTA countries:</p> <p data-bbox="460 427 927 451">Iceland, Liechtenstein, Norway, Switzerland</p> <p data-bbox="460 459 647 483">Other countries:</p> <p data-bbox="460 491 997 544">Colombia, El Salvador, Turkey, Ukraine, Australia, USA, Canada</p> <p data-bbox="460 552 972 604">FCC (Federal Communications Commission) notice</p> <ul style="list-style-type: none"> <li data-bbox="460 612 997 638">- This instrument fulfils part 15 of the FCC Rules <li data-bbox="460 646 964 699">- Its commissioning is subject to the following two conditions: <ol style="list-style-type: none"> <li data-bbox="497 707 997 759">1. this instrument must not cause any harmful interference and <li data-bbox="497 767 969 850">2. this instrument must be able to cope with interference, even if this has undesirable effects on operation. <p data-bbox="460 858 563 882">Changes</p> <p data-bbox="460 890 990 997">The FCC demands that the user be informed that any changes or modifications to the instrument that are not explicitly approved by Testo may void the user's right to use this instrument.</p>

Bluetooth® data 0554 0621

Feature	Values
Type designation	Stollmann E+V GmbH BlueMod+SR (August 2013)
Specification	4.0, Bluetooth® Classic / LowEnergy
Coverage	< 10 m / < 32,8 ft.
Product note	BlueMod+SR
Identification	B021281
Company	44784

Certification

Country	Comments				
EU + EFTA	<p>EU countries: Belgium (BE), Bulgaria (BG), Denmark (DK), Germany (DE), Estonia (EE), Finland (FI), France (FR), Greece (GR), Ireland (IE), Italy (IT), Latvia (LV), Lithuania (LT), Luxembourg (LU), Malta (MT), Netherlands (NL), Austria (AT), Poland (PL), Portugal (PT), Romania (RO), Sweden (SE), Slovakia (SK), Slovenia (SI), Spain (ES), Czech Republic (CZ), Hungary (HU), United Kingdom (GB), Republic of Cyprus (CY).</p> <p>EFTA countries: Iceland, Liechtenstein, Norway, Switzerland</p> <hr/> <p> You can find the EU declaration of conformity on the Testo website, www.testo.com, under the product-specific downloads.</p>				
Australia					
Turkey	Authorized				
India	Authorized				
Canada	Contains FCC ID: 4957A-MSR Product IC ID: -20160621 IC Warnings				
USA	Contains FCC ID: RFRMS Product FCC ID: WAF-20160621 FCC Warnings				
Korea	 MSIP-RMM-TE1-BTIRDA KCC Warning				
Japan	 <table border="1" data-bbox="529 1217 767 1289"> <tr> <td data-bbox="529 1217 565 1251">R</td> <td data-bbox="565 1217 767 1251">202-LSDO26</td> </tr> <tr> <td data-bbox="529 1251 565 1289">T</td> <td data-bbox="565 1251 767 1289">D 15-0014202</td> </tr> </table> Japan Information	R	202-LSDO26	T	D 15-0014202
R	202-LSDO26				
T	D 15-0014202				

IC Warnings

This instrument complies with Part 15C of the FCC Rules and Industry Canada RSS-210 (revision 8). Commissioning is subject to

the following two conditions:

- (1) This instrument must not cause any harmful interference and
- (2) this instrument must be able to cope with interference, even if this has undesirable effects on operation.

Cet appareil satisfait à la partie 15C des directives FCC et au standard Industrie Canada RSS-210 (révision 8). Sa mise en service est soumise aux deux conditions suivantes :

- (1) cet appareil ne doit causer aucune interférence dangereuse et
- (2) cet appareil doit supporter toute interférence, y compris des interférences qui provoquerait des opérations indésirables.

FCC Warnings

Information from the FCC (Federal Communications Commission)

For your own safety

Shielded cables should be used for a composite interface. This is to ensure continued protection against radio frequency interference.

FCC warning statement

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

Caution

Changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment. Shielded interface cable must be used in order to comply with the emission limits.

Warning

This device complies with Part 15 of the FCC Rules.

Operation is subject to the following two conditions:

- (1) this device may not cause harmful interference, and
- (2) this device must accept any interference received, including interference that may cause undesired operation.

KCC Warning

해당 무선 설비는 운용 중 전파혼신 가능성이 있음

Japan Information

当該機器には電波法に基づく、技術基準適合証明等を受けた特定無線設備を装着している。

Accessories/spare parts

Description	Item no.
Mains unit 5 V / 2.0 A with micro USB cable	0554 1107
Spare thermal paper, 6 rolls	0554 0568

4 Product description

4.1. Overview



- 1 Bluetooth®/infrared interface
- 2 Operating mode indicator
- 3 On/Off control key
- 4 Paper feed control key
- 5 Paper roll (printable side out)
- 6 Paper compartment cover
- 7 Magnetic holder (on back)

⚠ WARNING

⚠ Magnetic field
May be harmful to those with pacemakers.

> Keep a minimum distance of 10 cm between pacemaker and printer.

CAUTION

Magnetic field

Damage to other devices!

- > Keep a safe distance away from products that could be damaged by the effects of magnetism (e.g. monitors, computers or credit cards).



- 8 Micro USB connection for connecting to the mains unit or the PC.

4.2. Operating mode indicator

LED	Function
Steady green light	<ul style="list-style-type: none"> Standby mode Printing out print data
Flashing green light	<ul style="list-style-type: none"> Receiving print data
Steady red light	<ul style="list-style-type: none"> Not ready for operation Paper compartment cover open No paper inserted.
Flashing red light	<ul style="list-style-type: none"> Rechargeable battery low
Flashing red/green light	<ul style="list-style-type: none"> Rechargeable battery is charging
Off	<ul style="list-style-type: none"> Sleep mode, automatically activated 2 minutes after a button was last pressed

4.3. Control key functions

Control keys	Functions
	<ul style="list-style-type: none"> Switch on printer: press key < 2 s Switch off printer: press key > 2 s
	<ul style="list-style-type: none"> Short paper feed: press key < 1 s Paper feed for as long as the key is pressed: press key > 1 s

Control keys	Functions
 and 	<ul style="list-style-type: none">• Start a test print (printer is switched off): press keys < 1 s simultaneously• Bluetooth® module test (printer is off): press keys > 1 s simultaneously

5 First steps

5.1. Charging the rechargeable battery

The rechargeable battery can only be charged at an ambient temperature of ± 0 to $+35$ °C. If the rechargeable battery has been completely discharged, charging time at room temperature using the testo mains unit is approx. 6 h.

- > Connect battery charger to a mains socket and connect the printer.

i If stored for a prolonged period of time, the battery should be charged every once in a while because otherwise this may result in damage due to deep discharge.

Charging in the measuring instrument

1. Plug the mains unit instrument plug into the instrument's micro USB socket.
2. Plug the mains plug of the mains unit into a mains socket.
 - The charging process will start. The charging process will stop automatically when the battery is fully charged.

Battery care

- > Do not fully exhaust rechargeable batteries.

5.2. Inserting paper



1. Flip paper compartment cover up.
2. Insert paper roll, see illustration.
3. Close paper compartment cover.

6 Using the product

Printing data

i When first setting up a connection between a testo measuring instrument and the Bluetooth / IRDA printer, the initialization phase can last up to 30 seconds.

- ✓ Printer is switched on.
- > Start the print process from the instrument transmitting the data.
- Data is printed out (LED flashes green).



Testo SE & Co. KGaA

Celsiusstraße 2

79822 Titisee-Neustadt

Germany

Telefon: +49 7653 681-0

E-Mail: info@testo.de

Internet: www.testo.com